

2139 High Tech Drive State College, PA 16803 | USA 814-353-8000 | 800-676-6232 service@cannoninstrument.com cannoninstrument.com



## miniPV-HX viscometer Installation Guide

THANK YOU for your purchase of the miniPV-X automatic viscometer from CANNON Instrument Company. This guide is designed to provide information about the installation of the miniPV-X. This is designed to be used along with the miniPV-X Instruction & Operation Manual.

## 1. Unpack the miniPV-HX viscometer

Remove all components from the shipping container(s).



Do NOT lift the miniPV instrument by the modular side panels ("ears"). Always support the weight of the unit from the BASE.

- Remove all packing materials from the components.
- Verify reception of shipped materials by comparing equipment items with packing/parts lists. Report missing items to Cannon Instrument Company immediately.
- Inspect each component for signs of damage. Report damages to the shipper and Cannon Instrument Company immediately.
- Retain all packing materials until the instrument is connected and functioning properly.
- If any component(s) must be returned to Cannon Instrument Company, the damaged item(s) should be packaged in the original shipping container. Refer to the final page of this guide for instructions on returning defective equipment. Customers outside the United States should contact the local Cannon dealer for procedures on returning products to Cannon.
- Review operator safety information in the miniPV-X Instruction & Operation Manual
- Verify electrical specifications for MAINS power in the miniPV-X Instruction & Operation Manual.

**2.** Connect tubing and cables as indicated in the image. Cut tubing to length as necessary.

## **Tubing Detail**

A = 6 mm transparent tubing B = 1/8" transparent tubing

3. Fill bath vessel to the fill line with silicone fluid.

**4.** Install VISCPRO Software on the controlling PC using the installation CD or thumb drive.

5. Supply power to the miniPV instrument.

6. Load VISCPRO software.

- **7.** Set the desired control temperature.
- 8. Calibrate the instrument per manual instructions or confirm calibration data included with your installation CD.
- Enter sample ID information into VISCPRO and begin running samples.

